



Certificate of Calibration

Certificate Number: -L

Productivity Quality Inc.
15300 25th. Ave. North Suite 200
Plymouth, MN 55447
PH : 763-249-8130
FAX: 763-249-8150



Property of: MachinesUsed.com
Jill Hostad
2410 Niagra Lane N
Plymouth, MN 55447

Equipment Type: HAAS
Equipment Model: MINIMILL
Equipment Serial #: 1099641

Date of Inspection: 4/14/2023
Requested Due Date: 4/14/2024

Inspection Standard: ASME B5.54-2005
PQI Procedure: MTL_LLIN

Reference: Accreditation to ISO / IEC 17025 by
ANSI-ASQ National Accreditation Board / ANAB

Calibration Results

| | | Temp. Range | Fahrenheit | System | Maximum | Bidirectional | Mean | Expanded | Units |
|-------------|---------|-------------|------------|-----------|----------|---------------|-----------|-------------------------------------|-------|
| | | Start | Finish | Deviation | Reversal | Repeatability | Deviation | Uncertainty (+/- at 95% confidence) | |
| As Received | X Axis: | 74.40 | 74.40 | .000368 | .000048 | .000389 | .000343 | .000300 | IN |
| | Y Axis: | 73.60 | 73.60 | .000999 | .000095 | .000441 | .000995 | .000350 | IN |
| | Z Axis: | 75.10 | 75.10 | .000508 | .000108 | .000286 | .000460 | .000170 | IN |
| As Returned | X Axis: | 74.40 | 74.50 | .000133 | .000014 | .000257 | .000126 | .000300 | IN |
| | Y Axis: | 74.00 | 74.10 | .000552 | .000035 | .000297 | .000533 | .000350 | IN |
| | Z Axis: | 75.20 | 75.20 | .000284 | .000053 | .000142 | .000262 | .000170 | IN |

Calibration Equipment: Renishaw XM-60

Serial # : 4KGH31

Date of Last Calibration: 1/12/2023

Traceability Certificate # : 4KGH31-230112-00

Date Calibration Due: 1/31/2026

Traceability Certificate is traceable to NPL Standards. No assurances referencing the stability of results are made beyond the date of inspection.

Service Engineer: Adam Carlstrom

Approved By:

Any number of factors may cause a calibration item to drift out of calibration before the recommended interval has expired.

A Productivity Quality Inc. Certificate of Calibration may not be reproduced, except in full, without the written approval of Productivity Quality Inc.



Measurement Uncertainty

Certificate Number: -L

Productivity Quality Inc.
15300 25th. Ave. North Suite 200
Plymouth, MN 55447
PH : 763-249-8130
FAX: 763-249-8150



RENISHAW LASER INTERFEROMETER SERIAL #: 4KGH31

Standard Uncertainties

| Source of Uncertainty | Type | Value | Units | Dist. | Divisor | Constant | Length Depen. |
|---|------|----------|-------|--------|----------|----------|---------------|
| Operator Reproducibility | A | 0.000030 | MM | NORMAL | 1.000000 | 0.000030 | |
| Base Number From Laser Spec | B | 0.000025 | MM | RECT. | 1.732051 | 0.000014 | |
| Machine Repeatability (Highest Value Shown) | B | 0.007544 | MM | RECT. | 1.732051 | 0.004355 | |
| Laser Measurement System Accuracy | B | 0.700000 | PPM | RECT. | 1.732051 | | 0.404145 |
| Air Temp Changes Within Laser Path | B | 1.000000 | PPM | RECT. | 1.732051 | | 0.577350 |
| Air Temp Diff.: Sensor To Laser Path* | B | 0.046500 | PPM | RECT. | 1.732051 | | 0.026847 |
| Temp Sampling Error of ECU** | B | 0.046500 | PPM | RECT. | 1.732051 | | 0.026847 |

* OBSERVED TEMPERATURE DIFFERENCE: 0.05

** OBSERVED SAMPLING ERROR : 0.05

Uncertainty Summary Statement

| | X Axis | Y Axis | Z Axis | Units |
|---|----------|----------|----------|-------|
| 1. Combined Uncertainty for Constant Factors | 0.003769 | 0.004356 | 0.002083 | MM |
| 2. Expanded Uncertainty for Constant Factors (k=2) | 0.007538 | 0.008711 | 0.004165 | MM |
| 3. Combined Uncertainty for Dependent Factors | 0.705768 | 0.705768 | 0.705768 | PPM |
| 4. Expanded Uncertainty for Dependent Factors (k=2) | 1.411536 | 1.411536 | 1.411536 | PPM |

Stated Uncertainty

| | | | | |
|--|----------|----------|----------|-----|
| Expanded Uncertainty for Dependent Factors (k=2) | 1.411536 | 1.411536 | 1.411536 | PPM |
| Length of axis | 406.4 | 304.8 | 241.3 | MM |
| Expanded dependent uncertainty including axis length | 0.000574 | 0.000430 | 0.000341 | MM |
| Expanded uncertainty for constant factors | 0.007538 | 0.008711 | 0.004165 | MM |
| Expanded uncertainty for Single Axis (MM) | 0.007600 | 0.008800 | 0.004200 | MM |
| Expanded uncertainty for Single Axis (Inch) | .000300 | .000350 | .000170 | IN |

Expanded uncertainty Budget - Contains the single highest single axis uncertainty value

MM
0.008800
IN
.000350

The uncertainty is estimated using the principles of GUM, and a simple acceptance decision rule, in accordance to ASME B89.7.3.1-2001, applies for conformance verification. The uncertainty applies to this calibration only and not the measurement of customer parts. The uncertainty is the expanded uncertainty, with a coverage factor k=2, representing a confidence level of approximately 95%.



Calibration Summary

Certificate Number: -L

Productivity Quality Inc.
15300 25th. Ave. North Suite 200
Plymouth, MN 55447
PH : 763-249-8130
FAX: 763-249-8150



This "Summary" area is used by our Service Engineers to communicate concerns, special conditions, environmental instability, or any other information to expand upon your calibration service and equipment condition.

In many cases, it is difficult for us to make the decision whether or not your machine is "good". This is better determined by your company using the data we provide for you.

In addition, please feel free to call us with any questions you may have on our services and on the interpretation of the data we have provided. We would be pleased to provide assistance over the phone, through a personal visit or by forwarding additional documentation on the practices and standards we use in our calibration services.

Summary

The "As Returned" data reflects adjustments made to the backlash and 10e9 values. The compensation tables were not activated.

The "As Received" data reflects the original, as found, values for backlash and 10e9. Once the calibration was completed, these values were re-entered.

Your machine was found to be within what we consider reasonable limits and operational capability. There were no specific issues requiring immediate attention.